

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station: <u>Sheung Wun Yiu (AM1A)</u>	Operator: <u>Shum Kam Yuen</u>
Cal. Date: <u>31/5/2012</u>	Next Due Date: <u>31-Jul-12</u>
Equipment No.: <u>A-001-53T</u>	Serial No.: <u>10216</u>

Ambient Condition			
Temperature, Ta (K)	302.4	Pressure, Pa (mmHg)	755.9

Orifice Transfer Standard Information					
Serial No:	843	Slope, mc	2.00834	Intercept, bc	-0.02923
Last Calibration Date:	15-Nov-12	$mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$			
Next Calibration Date:	15-Nov-12				

Calibration of TSP Sampler					
Resistance Plate No.	Orifice			HVS Flow Recorder	
	DH (orifice), in. of water	$[DH \times (Pa/760) \times (298/Ta)]^{1/2}$	Qstd (m ³ /min) X-axis	Flow Recorder Reading (CFM)	Continuous Flow Recorder Reading IC (CFM) Y-axis
18	8.9	2.95	1.49	49.0	48.51
13	6.0	2.43	1.22	41.0	40.59
10	4.5	2.10	1.06	33.0	32.67
7	3.4	1.83	0.92	28.0	27.72
5	2.3	1.50	0.76	23.0	22.77

By Linear Regression of Y on X

Slope, mw = 36.7144 Intercept, bw = -5.5892

Correlation Coefficient* = 0.9932

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation	
From the TSP Field Calibration Curve, take Qstd = 1.30m ³ /min	
From the Regression Equation, the "Y" value according to	
$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$	
Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)] ^{1/2} =	<u>42.56</u>

Remarks: _____

QC Reviewer: K. H. SHEK Signature: Mike Date: 1 Jun 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station	Shan Tong New Village (AM2)	Operator:	Shum Kam Yuen
Cal. Date:	3-May-12	Next Due Date:	3-Jul-12
Equipment No.:	A-001-29T	Serial No.	10202

Ambient Condition			
Temperature, Ta (K)	300	Pressure, Pa (mmHg)	754.2

Orifice Transfer Standard Information					
Serial No:	843	Slope, mc	2.00834	Intercept, bc	-0.02923
Last Calibration Date:	15-Nov-11	$mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$			
Next Calibration Date:	15-Nov-12	$Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$			

Calibration of TSP Sampler					
Resistance Plate No.	Orifice			HVS Flow Recorder	
	DH (orifice), in. of water	$[DH \times (Pa/760) \times (298/Ta)]^{1/2}$	Qstd (m ³ /min) X-axis	Flow Recorder Reading (CFM)	Continuous Flow Recorder Reading IC (CFM) Y-axis
18	9.8	3.11	1.56	51.0	50.64
13	7.1	2.65	1.33	44.0	43.69
10	4.9	2.20	1.11	37.0	36.74
7	3.5	1.86	0.94	29.0	28.79
5	2.2	1.47	0.75	23.0	22.84

By Linear Regression of Y on X

Slope, mw = 34.7403 Intercept, bw = -2.9982

Correlation Coefficient* = 0.9944

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation	
From the TSP Field Calibration Curve, take Qstd = 1.30m ³ /min	
From the Regression Equation, the "Y" value according to	
$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$	
Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)] ^{1/2} =	<u>42.47</u>

Remarks: _____

QC Reviewer: K. H. SHEK Signature: Mike Date: 4 May 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station: Shan Tong New Village (AM2) Operator: Shum Kam Yuen
 Cal. Date: 03-Jul-12 Next Due Date: 03-Sep-12
 Equipment No.: A-001-29T Serial No.: 10202

Ambient Condition			
Temperature, Ta (K)	306	Pressure, Pa (mmHg)	752.7

Orifice Transfer Standard Information					
Serial No:	843	Slope, mc	2.00834	Intercept, bc	-0.02923
Last Calibration Date:	15-Nov-11	$mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$			
Next Calibration Date:	15-Nov-12	$Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$			

Calibration of TSP Sampler					
Resistance Plate No.	Orifice			HVS Flow Recorder	
	DH (orifice), in. of water	$[DH \times (Pa/760) \times (298/Ta)]^{1/2}$	Qstd (m ³ /min) X axis	Flow Recorder Reading (CFM)	Continuous Flow Recorder Reading IC (CFM) Y-axis
18	9.6	3.04	1.53	48.0	47.14
13	7.0	2.60	1.31	42.0	41.25
10	5.1	2.22	1.12	34.0	33.39
7	4.0	1.96	0.99	28.0	27.50
5	2.6	1.58	0.80	22.0	21.61

By Linear Regression of Y on X

Slope, mw = 36.3998

Intercept, bw = -7.7015

Correlation Coefficient* = 0.9918

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} = 40.34

Remarks: _____

QC Reviewer: 

Signature: 

Date: 5-Jul-12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station: <u>Riverain Bayside (AM3)</u>	Operator: <u>Shum Kam Yuen</u>
Cal. Date: <u>3-May-12</u>	Next Due Date: <u>3-Jul-12</u>
Equipment No.: <u>A-001-69T</u>	Serial No.: <u>716</u>

Ambient Condition			
Temperature, Ta (K)	303.8	Pressure, Pa (mmHg)	754.2

Orifice Transfer Standard Information					
Serial No:	843	Slope, mc	2.00834	Intercept, bc	-0.02923
Last Calibration Date:	15-Nov-11	$mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$			
Next Calibration Date:	15-Nov-12	$Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$			

Calibration of TSP Sampler					
Resistance Plate No.	Orifice			HVS Flow Recorder	
	DH (orifice), in. of water	[DH x (Pa/760) x (298/Ta)] ^{1/2}	Qstd (m ³ /min) X-axis	Flow Recorder Reading (CFM)	Continuous Flow Recorder Reading IC (CFM) Y-axis
18	9.6	3.06	1.54	51.0	50.32
13	7.4	2.68	1.35	42.0	41.44
10	5.6	2.33	1.18	37.0	36.51
7	4.4	2.07	1.05	33.0	32.56
5	2.3	1.50	0.76	21.0	20.72

By Linear Regression of Y on X

Slope, mw = 36.8237 Intercept, bw = -6.9183

Correlation Coefficient* = 0.9928

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]^{1/2} = 41.51

Remarks: _____

QC Reviewer: K. M. SHEK

Signature: Mike

Date: 4 May 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station Riverain Bayside (AM3) Operator: Shum Kam Yuen
 Cal. Date: 03-Jul-12 Next Due Date: 03-Sep-12
 Equipment No.: A-001-69T Serial No. 716

Ambient Condition			
Temperature, Ta (K)	306	Pressure, Pa (mmHg)	752.7

Orifice Transfer Standard Information					
Serial No:	843	Slope, mc	2.00834	Intercept, bc	-0.02923
Last Calibration Date:	15-Nov-11	$mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$			
Next Calibration Date:	15-Nov-12	$Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$			

Calibration of TSP Sampler					
Resistance Plate No.	Orifice			HVS Flow Recorder	
	DH (orifice), in. of water	$[DH \times (Pa/760) \times (298/Ta)]^{1/2}$	Qstd (m ³ /min) X axis	Flow Recorder Reading (CFM)	Continuous Flow Recorder Reading IC (CFM) Y-axis
18	8.4	2.85	1.43	44.0	43.21
13	7.6	2.71	1.36	42.0	41.25
10	6.1	2.43	1.22	38.0	37.32
7	4.8	2.15	1.09	32.0	31.43
5	3.0	1.70	0.86	24.0	23.57

By Linear Regression of Y on X

Slope, mw = 34.9045 Intercept, bw = -6.2805

Correlation Coefficient* = 0.9953

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m³/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = $(mw \times Qstd + bw) \times [(760 / Pa) \times (Ta / 298)]^{1/2} =$ 39.81

Remarks: _____

QC Reviewer: [Signature]

Signature: [Signature]

Date: 5 Jul 12

AECOM Asia Company Limited

TSP High Volume Sampler

Field Calibration Report

Station: 168 Shek Kwu Lung Village (AM4A)	Operator: Choi wing ho
Cal. Date: 31-May-12	Next Due Date: 31-Jul-12
Equipment No.: A-001-70T	Serial No.: 10273

Ambient Condition			
Temperature, Ta (K)	302.4	Pressure, Pa (mmHg)	755.9

Orifice Transfer Standard Information					
Serial No:	843	Slope, mc	2.00834	Intercept, bc	-0.02923
Last Calibration Date:	15-Nov-11	$mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$			
Next Calibration Date:	15-Nov-12	$Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$			

Calibration of TSP Sampler					
Resistance Plate No.	Orifice			HVS Flow Recorder	
	DH (orifice), in. of water	$[DH \times (Pa/760) \times (298/Ta)]^{1/2}$	Qstd (m ³ /min) X axis	Flow Recorder Reading (CFM)	Continuous Flow Recorder Reading IC (CFM) Y-axis
18	9.8	3.10	1.56	51.0	50.49
13	7.2	2.66	1.34	43.0	42.57
10	5.1	2.24	1.13	35.0	34.65
7	3.7	1.90	0.96	31.0	30.69
5	2.3	1.50	0.76	24.0	23.76

By Linear Regression of Y on X

Slope, mw = 33.3365 Intercept, bw = -1.8893

Correlation Coefficient* = 0.9962

*If Correlation Coefficient < 0.990, check and recalibrate.

Set Point Calculation	
From the TSP Field Calibration Curve, take Qstd = 1.30m ³ /min	
From the Regression Equation, the "Y" value according to	
$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$	
Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)] ^{1/2} =	41.87

Remarks: _____

QC Reviewer: K. H. SHEK Signature: Mike Date: 1 Jun 12



TISCH ENVIRONMENTAL, INC.
145 SOUTH MIAMI AVE.
VILLAGE OF CLEVELAND, OH 45002
513.467.9000
877.263.7610 TOLL FREE
513.467.9009 FAX
WWW.TISCH-ENV.COM

AIR POLLUTION MONITORING EQUIPMENT

ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5025A

Date - Nov 15, 2011 Rootsmeter S/N 0438320 Ta (K) - 294
Operator Tisch Orifice I.D. - 0843 Pa (mm) - 748.03

PLATE OR Run #	VOLUME START (m3)	VOLUME STOP (m3)	DIFF VOLUME (m3)	DIFF TIME (min)	METER DIFF Hg (mm)	ORFICE DIFF H2O (in.)
1	NA	NA	1.00	1.3810	3.2	2.00
2	NA	NA	1.00	0.9810	6.4	4.00
3	NA	NA	1.00	0.8760	7.8	5.00
4	NA	NA	1.00	0.8370	8.8	5.50
5	NA	NA	1.00	0.6890	12.7	8.00

DATA TABULATION

Vstd	(x axis) Qstd	(y axis)		Va	(x axis) Qa	(y axis)
0.9934	0.7193	1.4125		0.9957	0.7210	0.8866
0.9891	1.0083	1.9976		0.9915	1.0107	1.2538
0.9871	1.1269	2.2334		0.9895	1.1295	1.4018
0.9859	1.1779	2.3424		0.9882	1.1807	1.4703
0.9807	1.4233	2.8251		0.9830	1.4267	1.7732
Qstd slope (m) = 2.00834				Qa slope (m) = 1.25759		
intercept (b) = -0.02923				intercept (b) = -0.01835		
coefficient (r) = 0.99994				coefficient (r) = 0.99994		
y axis = SQRT[H2O(Pa/760) (298/Ta)]				y axis = SQRT[H2O(Ta/Pa)]		

CALCULATIONS

Vstd = Diff. Vol [(Pa-Diff. Hg)/760] (298/Ta)
Qstd = Vstd/Time

Va = Diff Vol [(Pa-Diff Hg)/Pa]
Qa = Va/Time

For subsequent flow rate calculations:

Qstd = 1/m{ [SQRT(H2O(Pa/760) (298/Ta))] - b}
Qa = 1/m{ [SQRT H2O(Ta/Pa)] - b}

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.07a
 Sensitivity Adjustment Scale Setting: 557 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 557 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 557 CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-06-12	13:30 - 14:30	27.9	63	0.04070	1628	27.13
2	02-06-12	14:30 - 15:30	27.9	63	0.04167	1669	27.82
3	02-06-12	15:30 - 16:30	28.2	64	0.04283	1713	28.55
4	02-06-12	16:30 - 17:30	28.1	63	0.04146	1655	27.58

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9951

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.08a
 Sensitivity Adjustment Scale Setting: 702 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 702 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 702 CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-07-12	13:30 - 14:30	28.9	73	0.04127	1545	25.75
2	02-07-12	14:30 - 15:30	29.0	73	0.04163	1566	26.10
3	02-07-12	15:30 - 16:30	29.0	73	0.04334	1630	27.17
4	02-07-12	16:30 - 17:30	29.1	74	0.04426	1665	27.74

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0016
 Correlation coefficient: 0.9952

Validity of Calibration Record: 1 July 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 3 July 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.09a
 Sensitivity Adjustment Scale Setting: 797 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No.: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 797 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 797 CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-06-12	13:30 - 14:30	27.9	63	0.04070	1626	27.10
2	02-06-12	14:30 - 15:30	27.9	63	0.04167	1667	27.78
3	02-06-12	15:30 - 16:30	28.2	64	0.04283	1708	28.47
4	02-06-12	16:30 - 17:30	28.1	63	0.04146	1659	27.65

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9949

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3
 Equipment No.: A.005.11a
 Sensitivity Adjustment Scale Setting: 799 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 799 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 799 CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-07-12	13:45 - 14:45	29.0	73	0.04152	1659	27.65
2	02-07-12	14:45 - 15:45	29.0	73	0.04194	1670	27.83
3	02-07-12	15:45 - 16:45	29.1	74	0.04318	1725	28.75
4	02-07-12	16:45 - 17:45	29.1	74	0.04443	1780	29.67

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9928

Validity of Calibration Record: 1 July 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 3 July 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
 Manufacturer/Brand: SIBATA
 Model No.: LD-3B
 Equipment No.: A.005.13a
 Sensitivity Adjustment Scale Setting: 643 CPM
 Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
 Venue: Cyberport (Pui Ying Secondary School)
 Model No.: Series 1400AB
 Serial No: Control: 140AB219899803
 Sensor: 1200C143659803 K₀: 12500
 Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 643 CPM
 Sensitivity Adjustment Scale Setting (After Calibration): 643 CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-06-12	13:30 - 14:30	27.9	63	0.04070	1623	27.05
2	02-06-12	14:30 - 15:30	27.9	63	0.04167	1663	27.72
3	02-06-12	15:30 - 16:30	28.2	64	0.04283	1771	28.52
4	02-06-12	16:30 - 17:30	28.1	63	0.04146	1656	27.60

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
 2. Total Count was logged by Laser Dust Monitor
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015
 Correlation coefficient: 0.9988

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type:	<i>Laser Dust Monitor</i>
Manufacturer/Brand:	<i>SIBATA</i>
Model No.:	<i>LD-3B</i>
Equipment No.:	<i>A.005.14a</i>
Sensitivity Adjustment Scale Setting:	<i>786 CPM</i>
Operator:	<i>Mike Shek (MSKM)</i>

Standard Equipment

Equipment:	<i>Rupprecht & Patashnick TEOM®</i>		
Venue:	<i>Cyberport (Pui Ying Secondary School)</i>		
Model No.:	<i>Series 1400AB</i>		
Serial No:	Control:	<i>140AB219899803</i>	
	Sensor:	<i>1200C143659803</i>	K _o : <i>12500</i>
Last Calibration Date*:	<i>5 May 2012</i>		

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration):	<u>786</u>	CPM
Sensitivity Adjustment Scale Setting (After Calibration):	786	CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-06-12	13:15 - 14:15	27.9	63	0.04073	1746	29.10
2	02-06-12	14:15 - 15:15	27.9	63	0.04154	1778	29.63
3	02-06-12	15:15 - 16:15	28.1	64	0.04269	1830	30.50
4	02-06-12	16:15 - 17:15	28.1	64	0.04136	1769	29.48

Note:

1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
2. Total Count was logged by Laser Dust Monitor
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor):	<u>0.0014</u>
Correlation coefficient:	0.9963

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung Signature: Date: 4 June 2012

EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor
Manufacturer/Brand: SIBATA
Model No.: LD-3B
Equipment No.: A.005.15a
Sensitivity Adjustment Scale Setting: 786 CPM

Operator: Mike Shek (MSKM)

Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®
Venue: Cyberport (Pui Ying Secondary School)
Model No.: Series 1400AB
Serial No: Control: 140AB219899803
Sensor: 1200C143659803 K_o: 12500
Last Calibration Date*: 5 May 2012

*Remarks: Recommended interval for hardware calibration is 1 year

Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 734 CPM
Sensitivity Adjustment Scale Setting (After Calibration): 734 CPM

Hour	Date (dd-mm-yy)	Time	Ambient Condition		Concentration ¹ (mg/m ³) Y-axis	Total Count ²	Count/ Minute ³ X-axis
			Temp (°C)	R.H. (%)			
1	02-06-12	13:15 - 14:15	27.9	63	0.04073	1748	29.13
2	02-06-12	14:15 - 15:15	27.9	63	0.04154	1780	29.67
3	02-06-12	15:15 - 16:15	28.1	64	0.04269	1826	30.43
4	02-06-12	16:15 - 17:15	28.1	64	0.04136	1773	29.55

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®
2. Total Count was logged by Laser Dust Monitor
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0014
Correlation coefficient: 0.9949

Validity of Calibration Record: 1 June 2013

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 4 June 2012



CERTIFICATE OF CALIBRATION

Certificate No.: 11CA1221 01-01

Page 1 of 2

Item tested

Description:	Sound Level Meter (Type 1)	Microphone	Preamp
Manufacturer:	Rion Co., Ltd.	Rion Co., Ltd.	Rion Co., Ltd.
Type/Model No.:	NL-31	UC-53A	NH-21
Serial/Equipment No.:	00320534 / N.007.02A	90526	03581
Adaptors used:	-	-	-

Item submitted by

Customer Name: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 21-Dec-2011

Date of test: 23-Dec-2011

Reference equipment used in the calibration

Description:	Model:	Serial No.	Expiry Date:	Traceable to:
Multi function sound calibrator	B&K 4226	2288444	09-May-2012	CIGISMEC
Signal generator	DS 360	33873	30-May-2012	CEPREI
Signal generator	DS 360	61227	30-May-2012	CEPREI

Ambient conditions

Temperature: (22 ± 1) °C
Relative humidity: (60 ± 10) %
Air pressure: (1000 ± 5) hPa

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of ±20%.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:


Huang Jian Min/Feng Jun Qi

Date: 16-Jan-2012

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 11CA0830 02 Page 1 of 2

Item tested

Description:	Sound Level Meter (Type 1)	Microphone	Preamp
Manufacturer:	Rion Co., Ltd.	Rion Co., Ltd.	Rion Co., Ltd.
Type/Model No.:	NL-31	UC-53A	NH-19
Serial/Equipment No.:	00320528 / N.007.03A	90565	75883
Adaptors used:	-	-	-

Item submitted by

Customer Name: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 30-Aug-2011

Date of test: 31-Aug-2011

Reference equipment used in the calibration

Description:	Model:	Serial No.	Expiry Date:	Traceable to:
Multi function sound calibrator	B&K 4226	2288444	09-May-2012	CIGISMEC
Signal generator	DS 360	33873	30-May-2012	CEPREI
Signal generator	DS 360	61227	30-May-2012	CEPREI

Ambient conditions

Temperature: $(23 \pm 1) ^\circ\text{C}$
Relative humidity: $(60 \pm 5) \%$
Air pressure: $(1000 \pm 5) \text{ hPa}$

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of $\pm 20\%$.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.


Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.


Actual Measurement data are documented on worksheets.

Approved Signatory:


Huang Jian-Aia/Feng Jun Qi

Date: 31-Aug-2011

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

Certificate No.: 12CA0321 01-04

Page: 1 of 2

Item tested

Description: Acoustical Calibrator (Class 1)
Manufacturer: Rion Co., Ltd.
Type/Model No.: NC-73
Serial/Equipment No.: 10186482 / N.004.09
Adaptors used: -

Item submitted by

Customer: AECOM ASIA CO., LTD.
Address of Customer: -
Request No.: -
Date of receipt: 21-Mar-2012

Date of test: 21-Mar-2012

Reference equipment used in the calibration

Description:	Model:	Serial No.	Expiry Date:	Traceable to:
Lab standard microphone	B&K 4180	2412857	18-May-2012	SCL
Preamplifier	B&K 2673	2239857	05-Jan-2013	CEPREI
Measuring amplifier	B&K 2610	2346941	29-Dec-2012	CEPREI
Signal generator	DS 360	61227	30-May-2012	CEPREI
Digital multi-meter	34401A	US36087050	16-Dec-2012	CEPREI
Audio analyzer	8903B	GB41300350	27-May-2012	CEPREI
Universal counter	53132A	MY40003662	30-May-2012	CEPREI

Ambient conditions

Temperature: 21 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 1005 ± 5 hPa

Test specifications

- The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

Test results

Details of the performed measurements are presented on page 2 of this certificate.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 23-Mar-2012

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.