

# AECOM Asia Company Limited

## TSP High Volume Sampler

### Field Calibration Report

|                |                       |                |           |
|----------------|-----------------------|----------------|-----------|
| Station        | Sheung Wun Yiu (AM1A) | Operator:      | Gary Choi |
| Cal. Date:     | 15-Jan-14             | Next Due Date: | 15-Mar-14 |
| Equipment No.: | A-001-53T             | Serial No.     | 10216     |

| Ambient Condition   |     |                     |       |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 289 | Pressure, Pa (mmHg) | 767.0 |

| Orifice Transfer Standard Information |           |                                                                    |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                          | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ |         |               |         |
| Next Calibration Date:                | 20-May-14 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 8.9                        | 3.04                                         | 1.55                              | 46.0                        | 46.93                                            |
| 13                         | 6.0                        | 2.50                                         | 1.27                              | 37.0                        | 37.74                                            |
| 10                         | 4.4                        | 2.14                                         | 1.09                              | 31.0                        | 31.62                                            |
| 7                          | 3.6                        | 1.94                                         | 0.98                              | 28.0                        | 28.56                                            |
| 5                          | 2.1                        | 1.48                                         | 0.75                              | 22.0                        | 22.44                                            |

By Linear Regression of Y on X

Slope, mw = 30.7427                      Intercept, bw = -1.2068

Correlation Coefficient\* = 0.9966

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation                                                                    |              |
|------------------------------------------------------------------------------------------|--------------|
| From the TSP Field Calibration Curve, take Qstd = 1.30m <sup>3</sup> /min                |              |
| From the Regression Equation, the "Y" value according to                                 |              |
| $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$                       |              |
| Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)] <sup>1/2</sup> = | <u>37.99</u> |

Remarks: \_\_\_\_\_

\_\_\_\_\_

QC Reviewer: WS CHAN                      Signature: [Signature]                      Date: 16/1/14

# AECOM Asia Company Limited

## TSP High Volume Sampler

### Field Calibration Report

|                |                       |                |           |
|----------------|-----------------------|----------------|-----------|
| Station        | Sheung Wun Yiu (AM1A) | Operator:      | Gary Choi |
| Cal. Date:     | 15-Mar-14             | Next Due Date: | 14-May-14 |
| Equipment No.: | A-001-53T             | Serial No.     | 10216     |

| Ambient Condition   |     |                     |       |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 289 | Pressure, Pa (mmHg) | 767.0 |

| Orifice Transfer Standard Information |           |                                                                    |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                          | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ |         |               |         |
| Next Calibration Date:                | 20-May-14 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 9.0                        | 3.06                                         | 1.56                              | 46.0                        | 46.93                                            |
| 13                         | 6.1                        | 2.52                                         | 1.28                              | 36.0                        | 36.72                                            |
| 10                         | 4.4                        | 2.14                                         | 1.09                              | 31.0                        | 31.62                                            |
| 7                          | 3.5                        | 1.91                                         | 0.97                              | 28.0                        | 28.56                                            |
| 5                          | 2.1                        | 1.48                                         | 0.75                              | 22.0                        | 22.44                                            |

By Linear Regression of Y on X

Slope, mw = 29.7418 Intercept, bw = -0.3150

Correlation Coefficient\* = 0.9931

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation                                                                            |              |
|--------------------------------------------------------------------------------------------------|--------------|
| From the TSP Field Calibration Curve, take Qstd = 1.30m <sup>3</sup> /min                        |              |
| From the Regression Equation, the "Y" value according to                                         |              |
| $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$                               |              |
| Therefore, Set Point; IC = ( mw x Qstd + bw ) x [ ( 760 / Pa ) x ( Ta / 298 ) ] <sup>1/2</sup> = | <u>37.59</u> |

Remarks: \_\_\_\_\_

QC Reviewer: Y1 Leung Signature: [Signature] Date: 15-3-14

# AECOM Asia Company Limited

## TSP High Volume Sampler

### Field Calibration Report

|                |                             |                |           |
|----------------|-----------------------------|----------------|-----------|
| Station        | Shan Tong New Village (AM2) | Operator:      | Gary Choi |
| Cal. Date:     | 15-Jan-14                   | Next Due Date: | 15-Mar-14 |
| Equipment No.: | A-001-29T                   | Serial No.     | 10202     |

| Ambient Condition   |     |                     |       |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 289 | Pressure, Pa (mmHg) | 767.0 |

| Orifice Transfer Standard Information |           |                                                                    |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                          | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ |         |               |         |
| Next Calibration Date:                | 20-May-14 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 10.0                       | 3.23                                         | 1.64                              | 46.0                        | 46.93                                            |
| 13                         | 7.5                        | 2.79                                         | 1.42                              | 41.0                        | 41.82                                            |
| 10                         | 5.5                        | 2.39                                         | 1.22                              | 35.0                        | 35.70                                            |
| 7                          | 3.8                        | 1.99                                         | 1.01                              | 28.0                        | 28.56                                            |
| 5                          | 2.5                        | 1.61                                         | 0.82                              | 23.0                        | 23.46                                            |

By Linear Regression of Y on X

Slope, mw = 29.0506 Intercept, bw = -0.2010

Correlation Coefficient\* = 0.9958

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation                                                                            |              |
|--------------------------------------------------------------------------------------------------|--------------|
| From the TSP Field Calibration Curve, take Qstd = 1.30m <sup>3</sup> /min                        |              |
| From the Regression Equation, the "Y" value according to                                         |              |
| $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$                               |              |
| Therefore, Set Point; IC = ( mw x Qstd + bw ) x [ ( 760 / Pa ) x ( Ta / 298 ) ] <sup>1/2</sup> = | <u>36.82</u> |

Remarks: \_\_\_\_\_

QC Reviewer: WS CHAN

Signature: [Signature]

Date: 16/1/14



# AECOM Asia Company Limited

## TSP High Volume Sampler

### Field Calibration Report

|                |                             |                |           |
|----------------|-----------------------------|----------------|-----------|
| Station        | Shan Tong New Village (AM2) | Operator:      | Gary Choi |
| Cal. Date:     | 15-Mar-14                   | Next Due Date: | 14-May-14 |
| Equipment No.: | A-001-29T                   | Serial No.     | 10202     |

| Ambient Condition   |     |                     |       |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 289 | Pressure, Pa (mmHg) | 767.0 |

| Orifice Transfer Standard Information |           |                                                                    |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                          | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ |         |               |         |
| Next Calibration Date:                | 20-May-14 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 10.0                       | 3.23                                         | 1.64                              | 48.0                        | 48.97                                            |
| 13                         | 7.5                        | 2.79                                         | 1.42                              | 42.0                        | 42.84                                            |
| 10                         | 5.6                        | 2.41                                         | 1.23                              | 35.0                        | 35.70                                            |
| 7                          | 3.7                        | 1.96                                         | 1.00                              | 28.0                        | 28.56                                            |
| 5                          | 2.5                        | 1.61                                         | 0.82                              | 24.0                        | 24.48                                            |

#### By Linear Regression of Y on X

Slope, mw = 30.3815 Intercept, bw = -0.9966

Correlation Coefficient\* = 0.9958

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation                                                                            |              |
|--------------------------------------------------------------------------------------------------|--------------|
| From the TSP Field Calibration Curve, take Qstd = 1.30m <sup>3</sup> /min                        |              |
| From the Regression Equation, the "Y" value according to                                         |              |
| $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$                               |              |
| Therefore, Set Point; IC = ( mw x Qstd + bw ) x [ ( 760 / Pa ) x ( Ta / 298 ) ] <sup>1/2</sup> = | <u>37.74</u> |

Remarks: \_\_\_\_\_

QC Reviewer: YT Leung Signature: [Signature] Date: 15-3-14

# AECOM Asia Company Limited

## TSP High Volume Sampler

### Field Calibration Report

|                                        |                                 |
|----------------------------------------|---------------------------------|
| Station: <u>Riverain Bayside (AM3)</u> | Operator: <u>Choi Wing Ho</u>   |
| Cal. Date: <u>18-Feb-14</u>            | Next Due Date: <u>18-May-14</u> |
| Equipment No.: <u>A-001-69T</u>        | Serial No. <u>716</u>           |

| Ambient Condition   |     |                     |       |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 290 | Pressure, Pa (mmHg) | 762.4 |

| Orifice Transfer Standard Information |           |                                                                    |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                          | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ |         |               |         |
| Next Calibration Date:                | 20-May-14 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 8.9                        | 3.03                                         | 1.54                              | 46.0                        | 46.70                                            |
| 13                         | 7.5                        | 2.78                                         | 1.42                              | 43.0                        | 43.66                                            |
| 10                         | 5.6                        | 2.40                                         | 1.22                              | 34.0                        | 34.52                                            |
| 7                          | 4.0                        | 2.03                                         | 1.03                              | 27.0                        | 27.41                                            |
| 5                          | 3.0                        | 1.76                                         | 0.89                              | 21.0                        | 21.32                                            |

By Linear Regression of Y on X

Slope, mw = 39.7791 Intercept, bw = -13.8329

Correlation Coefficient\* = 0.9949

\*If Correlation Coefficient < 0.990, check and recalibrate.

#### Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m<sup>3</sup>/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC =  $(mw \times Qstd + bw) \times [(760 / Pa) \times (Ta / 298)]^{1/2} =$  37.31

Remarks: \_\_\_\_\_

QC Reviewer: H. Sun

Signature: [Signature]

Date: 20-2-14



# **AECOM Asia Company Limited** **TSP High Volume Sampler** **Field Calibration Report**

|                                                 |                                 |
|-------------------------------------------------|---------------------------------|
| Station <u>168 Shek Kwu Lung Village (AM4A)</u> | Operator: <u>Gary Choi</u>      |
| Cal. Date: <u>15-Jan-14</u>                     | Next Due Date: <u>15-Mar-14</u> |
| Equipment No.: <u>A-001-70T</u>                 | Serial No. <u>10273</u>         |

| Ambient Condition   |     |                     |       |
|---------------------|-----|---------------------|-------|
| Temperature, Ta (K) | 289 | Pressure, Pa (mmHg) | 767.0 |

| Orifice Transfer Standard Information |           |                                                                                                                                      |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                                                                                            | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$ |         |               |         |
| Next Calibration Date:                | 20-May-14 |                                                                                                                                      |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 9.0                        | 3.06                                         | 1.56                              | 46.0                        | 46.93                                            |
| 13                         | 7.4                        | 2.78                                         | 1.41                              | 42.0                        | 42.84                                            |
| 10                         | 5.0                        | 2.28                                         | 1.16                              | 35.0                        | 35.70                                            |
| 7                          | 3.3                        | 1.85                                         | 0.94                              | 27.0                        | 27.54                                            |
| 5                          | 2.5                        | 1.61                                         | 0.82                              | 23.0                        | 23.46                                            |

By Linear Regression of Y on X

Slope, mw = 31.7150

Intercept, bw = -2.0528

Correlation Coefficient\* = 0.9964

\*If Correlation Coefficient < 0.990, check and recalibrate.

## Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 1.30m<sup>3</sup>/min

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)]<sup>1/2</sup> = 38.40

Remarks: \_\_\_\_\_

QC Reviewer: WS CHAN

Signature: [Signature]

Date: 16/1/14

# AECOM Asia Company Limited

## TSP High Volume Sampler

### Field Calibration Report

|                                           |                          |
|-------------------------------------------|--------------------------|
| Station: 168 Shek Kwu Lung Village (AM4A) | Operator: Gary Choi      |
| Cal. Date: 15-Mar-14                      | Next Due Date: 14-May-14 |
| Equipment No.: A-001-70T                  | Serial No.: 10273        |

| Ambient Condition   |       |                     |       |
|---------------------|-------|---------------------|-------|
| Temperature, Ta (K) | 289.4 | Pressure, Pa (mmHg) | 768.3 |

| Orifice Transfer Standard Information |           |                                                                    |         |               |         |
|---------------------------------------|-----------|--------------------------------------------------------------------|---------|---------------|---------|
| Serial No:                            | 988       | Slope, mc                                                          | 1.94727 | Intercept, bc | 0.02332 |
| Last Calibration Date:                | 20-May-13 | $mc \times Qstd + bc = [DH \times (Pa/760) \times (298/Ta)]^{1/2}$ |         |               |         |
| Next Calibration Date:                | 20-May-14 | $Qstd = \{[DH \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |         |               |         |

| Calibration of TSP Sampler |                            |                                              |                                   |                             |                                                  |
|----------------------------|----------------------------|----------------------------------------------|-----------------------------------|-----------------------------|--------------------------------------------------|
| Resistance Plate No.       | Orifice                    |                                              |                                   | HVS Flow Recorder           |                                                  |
|                            | DH (orifice), in. of water | $[DH \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (m <sup>3</sup> /min) X-axis | Flow Recorder Reading (CFM) | Continuous Flow Recorder Reading IC (CFM) Y-axis |
| 18                         | 8.8                        | 3.03                                         | 1.54                              | 46.0                        | 46.93                                            |
| 13                         | 7.4                        | 2.78                                         | 1.41                              | 42.0                        | 42.85                                            |
| 10                         | 5.1                        | 2.30                                         | 1.17                              | 36.0                        | 36.73                                            |
| 7                          | 3.2                        | 1.83                                         | 0.93                              | 28.0                        | 28.57                                            |
| 5                          | 2.5                        | 1.61                                         | 0.82                              | 23.0                        | 23.47                                            |

#### By Linear Regression of Y on X

Slope, mw = 31.3458 Intercept, bw = -1.0821

Correlation Coefficient\* = 0.9922

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation                                                                    |              |
|------------------------------------------------------------------------------------------|--------------|
| From the TSP Field Calibration Curve, take Qstd = 1.30m <sup>3</sup> /min                |              |
| From the Regression Equation, the "Y" value according to                                 |              |
| $mw \times Qstd + bw = IC \times [(Pa/760) \times (298/Ta)]^{1/2}$                       |              |
| Therefore, Set Point; IC = (mw x Qstd + bw) x [(760 / Pa) x (Ta / 298)] <sup>1/2</sup> = | <u>38.88</u> |

Remarks: \_\_\_\_\_

QC Reviewer: Y.T. Wong Signature: [Signature] Date: 15-3-14





TISCH ENVIRONMENTAL, INC.  
145 SOUTH MIAMI AVE.  
VILLAGE OF CLEVELAND, OH 45002  
513.467.9000  
877.263.7610 TOLL FREE  
513.467.9009 FAX  
WWW.TISCH-ENV.COM

# AIR POLLUTION MONITORING EQUIPMENT

## ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5025A

Date - May 20, 2013 Rootsmeter S/N 0438320 Ta (K) - 297  
Operator Tisch Orifice I.D. - 0988 Pa (mm) - 751.84

| PLATE<br>OR<br>Run # | VOLUME<br>START<br>(m3) | VOLUME<br>STOP<br>(m3) | DIFF<br>VOLUME<br>(m3) | DIFF<br>TIME<br>(min) | METER              | ORFICE               |
|----------------------|-------------------------|------------------------|------------------------|-----------------------|--------------------|----------------------|
|                      |                         |                        |                        |                       | DIFF<br>Hg<br>(mm) | DIFF<br>H2O<br>(in.) |
| 1                    | NA                      | NA                     | 1.00                   | 1.3900                | 3.2                | 2.00                 |
| 2                    | NA                      | NA                     | 1.00                   | 0.9720                | 6.4                | 4.00                 |
| 3                    | NA                      | NA                     | 1.00                   | 0.8670                | 7.9                | 5.00                 |
| 4                    | NA                      | NA                     | 1.00                   | 0.8270                | 8.7                | 5.50                 |
| 5                    | NA                      | NA                     | 1.00                   | 0.6800                | 12.6               | 8.00                 |

## DATA TABULATION

| Vstd                                | (x axis)<br>Qstd | (y axis) | Va                        | (x axis)<br>Qa | (y axis) |
|-------------------------------------|------------------|----------|---------------------------|----------------|----------|
| 0.9884                              | 0.7110           | 1.4090   | 0.9957                    | 0.7163         | 0.8889   |
| 0.9842                              | 1.0125           | 1.9926   | 0.9915                    | 1.0201         | 1.2570   |
| 0.9821                              | 1.1327           | 2.2278   | 0.9894                    | 1.1412         | 1.4054   |
| 0.9811                              | 1.1863           | 2.3365   | 0.9884                    | 1.1952         | 1.4740   |
| 0.9759                              | 1.4352           | 2.8179   | 0.9832                    | 1.4459         | 1.7777   |
| Qstd slope (m) = 1.94727            |                  |          | Qa slope (m) = 1.21935    |                |          |
| intercept (b) = 0.02332             |                  |          | intercept (b) = 0.01471   |                |          |
| coefficient (r) = 0.99998           |                  |          | coefficient (r) = 0.99998 |                |          |
| y axis = SQRT[H2O(Pa/760) (298/Ta)] |                  |          | y axis = SQRT[H2O(Ta/Pa)] |                |          |

## CALCULATIONS

$$Vstd = \text{Diff. Vol}[(Pa - \text{Diff. Hg})/760] (298/Ta)$$

$$Qstd = Vstd/\text{Time}$$

$$Va = \text{Diff Vol} [(Pa - \text{Diff Hg})/Pa]$$

$$Qa = Va/\text{Time}$$

For subsequent flow rate calculations:

$$Qstd = 1/m\{[\text{SQRT}(H2O(Pa/760) (298/Ta))] - b\}$$

$$Qa = 1/m\{[\text{SQRT } H2O(Ta/Pa)] - b\}$$



## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
 Manufacturer/Brand: SIBATA  
 Model No.: LD-3  
 Equipment No.: A.005.07a  
 Sensitivity Adjustment Scale Setting: 557 CPM  
 Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
 Venue: Cyberport (Pui Ying Secondary School)  
 Model No.: Series 1400AB  
 Serial No: Control: 140AB219899803  
 Sensor: 1200C143659803 K<sub>o</sub>: 12500  
 Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 557 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration): 557 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|-------------------|-------------|--------------------------------------------------------------|--------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)      | R.H.<br>(%) |                                                              |                          |                                         |
| 1    | 18-05-13           | 12:30 - 13:30 | 28.1              | 78          | 0.04714                                                      | 1887                     | 31.45                                   |
| 2    | 18-05-13           | 13:30 - 14:30 | 28.1              | 78          | 0.04932                                                      | 1970                     | 32.83                                   |
| 3    | 18-05-13           | 14:30 - 15:30 | 28.2              | 77          | 0.05156                                                      | 2056                     | 34.27                                   |
| 4    | 18-05-13           | 15:30 - 16:30 | 28.1              | 78          | 0.05083                                                      | 2026                     | 33.77                                   |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
 2. Total Count was logged by Laser Dust Monitor  
 3. Count/minute was calculated by (Total Count/60)


By Linear Regression of Y or X

Slope (K-factor): 0.0015  
 Correlation coefficient: 0.9978

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 20 May 2013

## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
Manufacturer/Brand: SIBATA  
Model No.: LD-3  
Equipment No.: A.005.08a  
Sensitivity Adjustment Scale Setting: 702 CPM  
  
Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
Venue: Cyberport (Pui Ying Secondary School)  
Model No.: Series 1400AB  
Serial No: Control: 140AB219899803  
Sensor: 1200C143659803 K<sub>0</sub>: 12500  
Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 702 CPM  
Sensitivity Adjustment Scale Setting (After Calibration): 702 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient<br>Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total<br>Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)         | R.H.<br>(%) |                                                              |                             |                                         |
| 1    | 18-05-13           | 12:30 - 13:30 | 28.1                 | 78          | 0.04714                                                      | 1764                        | 29.40                                   |
| 2    | 18-05-13           | 13:30 - 14:30 | 28.1                 | 78          | 0.04932                                                      | 1846                        | 30.77                                   |
| 3    | 18-05-13           | 14:30 - 15:30 | 28.2                 | 77          | 0.05156                                                      | 1935                        | 32.25                                   |
| 4    | 18-05-13           | 15:30 - 16:30 | 28.1                 | 78          | 0.05083                                                      | 1899                        | 31.65                                   |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
2. Total Count was logged by Laser Dust Monitor  
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0016  
Correlation coefficient: 0.9976

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013



## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
Manufacturer/Brand: SIBATA  
Model No.: LD-3  
Equipment No.: A.005.09a  
Sensitivity Adjustment Scale Setting: 797 CPM  
  
Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
Venue: Cyberport (Pui Ying Secondary School)  
Model No.: Series 1400AB  
Serial No: Control: 140AB219899803  
Sensor: 1200C143659803 K<sub>0</sub>: 12500  
Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 797 CPM  
Sensitivity Adjustment Scale Setting (After Calibration): 797 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|-------------------|-------------|--------------------------------------------------------------|--------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)      | R.H.<br>(%) |                                                              |                          |                                         |
| 1    | 18-05-13           | 12:30 - 13:30 | 28.1              | 78          | 0.04714                                                      | 1885                     | 31.42                                   |
| 2    | 18-05-13           | 13:30 - 14:30 | 28.1              | 78          | 0.04932                                                      | 1965                     | 32.75                                   |
| 3    | 18-05-13           | 14:30 - 15:30 | 28.2              | 77          | 0.05156                                                      | 2059                     | 34.32                                   |
| 4    | 18-05-13           | 15:30 - 16:30 | 28.1              | 78          | 0.05083                                                      | 2024                     | 33.73                                   |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
2. Total Count was logged by Laser Dust Monitor  
3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015  
Correlation coefficient: 0.9973

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
 Manufacturer/Brand: SIBATA  
 Model No.: LD-3  
 Equipment No.: A.005.10a  
 Sensitivity Adjustment Scale Setting: 753 CPM  
 Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
 Venue: Cyberport (Pui Ying Secondary School)  
 Model No.: Series 1400AB  
 Serial No: Control: 140AB219899803  
 Sensor: 1200C143659803 K<sub>0</sub>: 12500  
 Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 753 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration): 753 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient<br>Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total<br>Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)         | R.H.<br>(%) |                                                              |                             |                                         |
| 1    | 18-05-13           | 12:30 - 13:30 | 28.1                 | 78          | 0.04714                                                      | 1886                        | 31.43                                   |
| 2    | 18-05-13           | 13:30 - 14:30 | 28.1                 | 78          | 0.04932                                                      | 1968                        | 32.80                                   |
| 3    | 18-05-13           | 14:30 - 15:30 | 28.2                 | 77          | 0.05156                                                      | 2061                        | 34.35                                   |
| 4    | 18-05-13           | 15:30 - 16:30 | 28.1                 | 78          | 0.05083                                                      | 2026                        | 33.77                                   |


Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
 2. Total Count was logged by Laser Dust Monitor  
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015  
 Correlation coefficient: 0.9983

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung Signature:  Date: 20 May 2013

## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
 Manufacturer/Brand: SIBATA  
 Model No.: LD-3  
 Equipment No.: A.005.11a  
 Sensitivity Adjustment Scale Setting: 799 CPM  
 Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
 Venue: Cyberport (Pui Ying Secondary School)  
 Model No.: Series 1400AB  
 Serial No: Control: 140AB219899803  
 Sensor: 1200C143659803 K<sub>0</sub>: 12500  
 Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 799 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration): 799 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|-------------------|-------------|--------------------------------------------------------------|--------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)      | R.H.<br>(%) |                                                              |                          |                                         |
| 1    | 18-05-13           | 12:15 - 13:15 | 28.1              | 78          | 0.04685                                                      | 1871                     | 31.18                                   |
| 2    | 18-05-13           | 13:15 - 14:15 | 28.1              | 78          | 0.04941                                                      | 1979                     | 32.98                                   |
| 3    | 18-05-13           | 14:15 - 15:15 | 28.2              | 77          | 0.05127                                                      | 2055                     | 34.25                                   |
| 4    | 18-05-13           | 15:15 - 16:15 | 28.1              | 78          | 0.05060                                                      | 2021                     | 33.68                                   |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
 2. Total Count was logged by Laser Dust Monitor  
 3. Count/minute was calculated by (Total Count/60)

By Linear Regression of Y or X

Slope (K-factor): 0.0015  
 Correlation coefficient: 0.9976

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 20 May 2013



## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
 Manufacturer/Brand: SIBATA  
 Model No.: LD-3B  
 Equipment No.: A.005.13a  
 Sensitivity Adjustment Scale Setting: 643 CPM  
 Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
 Venue: Cyberport (Pui Ying Secondary School)  
 Model No.: Series 1400AB  
 Serial No: Control: 140AB219899803  
 Sensor: 1200C143659803 K<sub>0</sub>: 12500  
 Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 643 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration): 643 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient<br>Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total<br>Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|----------------------|-------------|--------------------------------------------------------------|-----------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)         | R.H.<br>(%) |                                                              |                             |                                         |
| 1    | 18-05-13           | 12:15 - 13:15 | 28.1                 | 78          | 0.04685                                                      | 1867                        | 31.12                                   |
| 2    | 18-05-13           | 13:15 - 14:15 | 28.1                 | 78          | 0.04941                                                      | 1975                        | 32.92                                   |
| 3    | 18-05-13           | 14:15 - 15:15 | 28.2                 | 77          | 0.05127                                                      | 2048                        | 34.13                                   |
| 4    | 18-05-13           | 15:15 - 16:15 | 28.1                 | 78          | 0.05060                                                      | 2017                        | 33.62                                   |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
 2. Total Count was logged by Laser Dust Monitor  
 3. Count/minute was calculated by (Total Count/60)


By Linear Regression of Y or X

Slope (K-factor): 0.0015  
 Correlation coefficient: 0.9986

Validity of Calibration Record: 17 May 2014

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 20 May 2013

## EQUIPMENT CALIBRATION RECORD

Type: Laser Dust Monitor  
 Manufacturer/Brand: SIBATA  
 Model No.: LD-3B  
 Equipment No.: A.005.16a  
 Sensitivity Adjustment Scale Setting: 521 CPM  
 Operator: Mike Shek (MSKM)

### Standard Equipment

Equipment: Rupprecht & Patashnick TEOM®  
 Venue: Cyberport (Pui Ying Secondary School)  
 Model No.: Series 1400AB  
 Serial No: Control: 140AB219899803  
 Sensor: 1200C143659803 K<sub>o</sub>: 12500  
 Last Calibration Date\*: 18 May 2013

\*Remarks: Recommended interval for hardware calibration is 1 year

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration): 521 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration): 521 CPM

| Hour | Date<br>(dd-mm-yy) | Time          | Ambient Condition |             | Concentration <sup>1</sup><br>(mg/m <sup>3</sup> )<br>Y-axis | Total Count <sup>2</sup> | Count/<br>Minute <sup>3</sup><br>X-axis |
|------|--------------------|---------------|-------------------|-------------|--------------------------------------------------------------|--------------------------|-----------------------------------------|
|      |                    |               | Temp<br>(°C)      | R.H.<br>(%) |                                                              |                          |                                         |
| 1    | 27-07-13           | 11:00 - 12:00 | 27.3              | 75          | 0.04734                                                      | 1893                     | 31.55                                   |
| 2    | 27-07-13           | 12:00 - 13:00 | 27.3              | 75          | 0.04789                                                      | 1915                     | 31.92                                   |
| 3    | 27-07-13           | 13:00 - 14:00 | 27.4              | 74          | 0.04953                                                      | 1976                     | 32.93                                   |
| 4    | 27-07-13           | 14:00 - 15:00 | 27.4              | 75          | 0.04867                                                      | 1949                     | 32.48                                   |

Note: 1. Monitoring data was measured by Rupprecht & Patashnick TEOM®  
 2. Total Count was logged by Laser Dust Monitor  
 3. Count/minute was calculated by (Total Count/60)


By Linear Regression of Y or X

Slope (K-factor): 0.0015  
 Correlation coefficient: 0.9934

Validity of Calibration Record: 26 July 2014

Remarks:

QC Reviewer: YW Fung

Signature: 

Date: 29 July 2013



## CERTIFICATE OF CALIBRATION

Certificate No.: 13CA1107 01-02

Page: 1 of 2

### Item tested

Description: Acoustical Calibrator (Class 1)  
Manufacturer: Rion Co., Ltd.  
Type/Model No.: NC-73  
Serial/Equipment No.: 10307223 / N.004.08  
Adaptors used: -

### Item submitted by

Customer: AECOM ASIA CO., LTD.  
Address of Customer: -  
Request No.: -  
Date of receipt: 07-Nov-2013

Date of test: 08-Nov-2013

### Reference equipment used in the calibration

| Description:            | Model:   | Serial No. | Expiry Date: | Traceable to: |
|-------------------------|----------|------------|--------------|---------------|
| Lab standard microphone | B&K 4180 | 2341427    | 17-Apr-2014  | SCL           |
| Preamplifier            | B&K 2673 | 2239857    | 16-Apr-2014  | CEPREI        |
| Measuring amplifier     | B&K 2610 | 2346941    | 24-Apr-2014  | CEPREI        |
| Signal generator        | DS 360   | 61227      | 15-Apr-2014  | CEPREI        |
| Digital multi-meter     | 34401A   | US36087050 | 10-Dec-2013  | CEPREI        |
| Audio analyzer          | 8903B    | GB41300350 | 15-Apr-2014  | CEPREI        |
| Universal counter       | 53132A   | MY40003662 | 15-Apr-2014  | CEPREI        |

### Ambient conditions

Temperature:  $22 \pm 1$  °C  
Relative humidity:  $60 \pm 10$  %  
Air pressure:  $1000 \pm 10$  hPa

### Test specifications

- The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

### Test results

This is to certify that the sound calibrator conforms to the requirements of annex B of IEC 60942: 1997 for the conditions under which the test was performed. This does not imply that the sound calibrator meets IEC 60942 under any other conditions.

Details of the performed measurements are presented on **page 2** of this certificate.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 11-Nov-2013

Company Chop:



**Comments:** The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.





## CERTIFICATE OF CALIBRATION

Certificate No.: 13CA0325 01-03

Page: 1 of 2

### Item tested

Description: Acoustical Calibrator (Class 1)  
Manufacturer: Rion Co., Ltd.  
Type/Model No.: NC-73  
Serial/Equipment No.: 10186482 / N.004.09  
Adaptors used: -

### Item submitted by

Customer: AECOM ASIA CO., LTD.  
Address of Customer: -  
Request No.: -  
Date of receipt: 25-Mar-2013

Date of test: 26-Mar-2013

### Reference equipment used in the calibration

| Description:            | Model:   | Serial No. | Expiry Date: | Traceable to: |
|-------------------------|----------|------------|--------------|---------------|
| Lab standard microphone | B&K 4180 | 2412857    | 29-May-2013  | SCL           |
| Preamplifier            | B&K 2673 | 2239857    | 17-Dec-2013  | CEPREI        |
| Measuring amplifier     | B&K 2610 | 2346941    | 17-Dec-2013  | CEPREI        |
| Signal generator        | DS 360   | 61227      | 29-May-2013  | CEPREI        |
| Digital multi-meter     | 34401A   | US36087050 | 10-Dec-2013  | CEPREI        |
| Audio analyzer          | 8903B    | GB41300350 | 29-May-2013  | CEPREI        |
| Universal counter       | 53132A   | MY40003662 | 29-May-2013  | CEPREI        |

### Ambient conditions

Temperature:  $22 \pm 1$  °C  
Relative humidity:  $60 \pm 10$  %  
Air pressure:  $1000 \pm 10$  hPa

### Test specifications

- 1, The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- 2, The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- 3, The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

### Test results

This is to certify that the sound calibrator conforms to the requirements of annex B of IEC 60942: 1997 for the conditions under which the test was performed. This does not imply that the sound calibrator meets IEC 60942 under any other conditions.

Details of the performed measurements are presented on page 2 of this certificate.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 26-Mar-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



## CERTIFICATE OF CALIBRATION

Certificate No.: 13CA1107 01-01

Page 1 of 2

### Item tested

|                       |                            |                  |
|-----------------------|----------------------------|------------------|
| Description:          | Sound Level Meter (Type 1) | , Microphone     |
| Manufacturer:         | Rion Co., Ltd.             | , Rion Co., Ltd. |
| Type/Model No.:       | NL-31                      | , UC-53A         |
| Serial/Equipment No.: | 00320528 / N.007.03A       | , 90565          |
| Adaptors used:        | -                          | , -              |

### Item submitted by

Customer Name: AECOM ASIA CO., LTD.  
Address of Customer: -  
Request No.: -  
Date of receipt: 07-Nov-2013

Date of test: 08-Nov-2013

### Reference equipment used in the calibration

| Description:                    | Model:   | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444    | 22-Jun-2014  | CIGISMEC      |
| Signal generator                | DS 360   | 33873      | 15-Apr-2014  | CEPREI        |
| Signal generator                | DS 360   | 61227      | 15-Apr-2014  | CEPREI        |

### Ambient conditions

Temperature:  $22 \pm 1^\circ\text{C}$   
Relative humidity:  $60 \pm 10\%$   
Air pressure:  $1000 \pm 10\text{ hPa}$

### Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of  $\pm 20\%$ .
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

### Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:

  
Huang Jian Min/Feng Jun Qi

Date: 11-Nov-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.





綜合試驗有限公司  
SOILS & MATERIALS ENGINEERING CO., LTD.

G/F, 9/F, 12/F, 13/F. & 20/F., Leader Centre, 37 Wong Chuk Hang Road, Aberdeen, Hong Kong.  
香港黃竹坑道37號利達中心地下, 9樓, 12樓, 13樓及20樓  
E-mail: smec@cigismec.com Website: www.cigismec.com

Tel : (852) 2873 6860  
Fax : (852) 2555 7533



## CERTIFICATE OF CALIBRATION

Certificate No.: 13CA0325 01-01

Page 1 of 2

### Item tested

|                       |                            |            |
|-----------------------|----------------------------|------------|
| Description:          | Sound Level Meter (Type 1) | Microphone |
| Manufacturer:         | B & K                      | B & K      |
| Type/Model No.:       | 2238                       | 4188       |
| Serial/Equipment No.: | 2285692 1609.04            | 2250420    |
| Adaptors used:        | -                          | -          |

### Item submitted by

Customer Name: AECOM ASIA CO., LTD.  
Address of Customer: -  
Request No.: -  
Date of receipt: 25-Mar-2013

Date of test: 26-Mar-2013

### Reference equipment used in the calibration

| Description:                    | Model:   | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444    | 22-Jun-2013  | CIGISMEC      |
| Signal generator                | DS 360   | 33873      | 29-May-2013  | CEPREI        |
| Signal generator                | DS 360   | 61227      | 29-May-2013  | CEPREI        |

### Ambient conditions

Temperature:  $22 \pm 1^\circ\text{C}$   
Relative humidity:  $60 \pm 10\%$   
Air pressure:  $1000 \pm 10\text{ hPa}$

### Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of  $\pm 20\%$ .
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure response of the Sound Level Meter.

### Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:

Huang Jian Min/Feng Jun Qi

Date: 26-Mar-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.





## CERTIFICATE OF CALIBRATION

Certificate No.: 13CA0617 01-01

Page 1 of 2

### Item tested

|                       |                            |   |            |
|-----------------------|----------------------------|---|------------|
| Description:          | Sound Level Meter (Type 1) | , | Microphone |
| Manufacturer:         | B & K                      | , | B & K      |
| Type/Model No.:       | 2238                       | , | 4188       |
| Serial/Equipment No.: | 2800927 / N.009.06         | , | 2791211    |
| Adaptors used:        | -                          | , | -          |

### Item submitted by

Customer Name: AECOM ASIA CO. LTD.  
Address of Customer: -  
Request No.: -  
Date of receipt: 17-Jun-2013

Date of test: 18-Jun-2013

### Reference equipment used in the calibration

| Description:                    | Model:   | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444    | 22-Jun-2013  | CIGISMEC      |
| Signal generator                | DS 360   | 33873      | 15-Apr-2014  | CEPREI        |
| Signal generator                | DS 360   | 61227      | 15-Apr-2014  | CEPREI        |

### Ambient conditions

Temperature:  $22 \pm 1$  °C  
Relative humidity:  $60 \pm 10$  %  
Air pressure:  $1000 \pm 10$  hPa

### Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of  $\pm 20\%$ .
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure response of the Sound Level Meter.

### Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.


Actual Measurement data are documented on worksheets.

Approved Signatory:

Huang Jian Min / Feng Jun Qi

Date: 18-Jun-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



## CERTIFICATE OF CALIBRATION

Certificate No.: 13CA0305 01-01

Page 1 of 2

### Item tested

|                       |                            |              |
|-----------------------|----------------------------|--------------|
| Description:          | Sound Level Meter (Type 1) | , Microphone |
| Manufacturer:         | B & K                      | , B & K      |
| Type/Model No.:       | 2250-L                     | , 4950       |
| Serial/Equipment No.: | 2681366 (N.O.II.OI)        | , 2665582    |
| Adaptors used:        | -                          | , -          |

### Item submitted by

Customer Name: AECOM ASIA CO LIMITED  
Address of Customer: -  
Request No.: -  
Date of receipt: 05-Mar-2013

Date of test: 05-Mar-2013

### Reference equipment used in the calibration

| Description:                    | Model:   | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444    | 23-May-2013  | CIGISMEC      |
| Signal generator                | DS 360   | 33873      | 29-May-2013  | CEPREI        |
| Signal generator                | DS 360   | 61227      | 29-May-2013  | CEPREI        |

### Ambient conditions

Temperature:  $21 \pm 1$  °C  
Relative humidity:  $60 \pm 10$  %  
Air pressure:  $1000 \pm 10$  hPa

### Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of  $\pm 20\%$ .
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

### Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:

  
Huang Jian Min/Feng Jun Qi

Date: 05-Mar-2013

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.





## CERTIFICATE OF CALIBRATION

Certificate No.: 14CA0305 06-02

Page 1 of 2

### Item tested

|                       |                            |            |
|-----------------------|----------------------------|------------|
| Description:          | Sound Level Meter (Type 1) | Microphone |
| Manufacturer:         | B & K                      | B & K      |
| Type/Model No.:       | 2250                       | 4950       |
| Serial/Equipment No.: | 2681366 N.011.01           | 2665582    |
| Adaptors used:        | -                          | -          |

### Item submitted by

Customer Name: AECOM ASIA CO. LTD.  
Address of Customer: -  
Request No.: -  
Date of receipt: 05-Mar-2014

Date of test: 07-Mar-2014

### Reference equipment used in the calibration

| Description:                    | Model:   | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444    | 22-Jun-2014  | CIGISMEC      |
| Signal generator                | DS 360   | 33873      | 15-Apr-2014  | CEPREI        |
| Signal generator                | DS 360   | 61227      | 15-Apr-2014  | CEPREI        |

### Ambient conditions

Temperature:  $22 \pm 1$  °C  
Relative humidity:  $60 \pm 10$  %  
Air pressure:  $1000 \pm 10$  hPa

### Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of  $\pm 20\%$ .
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responsiveness of the Sound Level Meter.

### Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:

  
Huang Jian Min / Feng Jun Qi

Date: 12-Mar-2014

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.